

June 18, 2021

Dear Sir/Madam,

**Subject:** Notification regarding "Information on Cybersecurity of Medical devices"  
for Adamonitor SC

Otsuka Electronics Co., Ltd has issued this notification letter to announce that in accordance with the guidelines "Principles and Practices for Medical Device Cybersecurity (2020.3.18)", has published the information regarding the End of Life (EOL) and End of Support (EOS) of medical devices listed below on our official website.

Product Name: Adamonitor SC, Adastand

Note: Adastand is Accessory for Adamonitor SC.

The definitions of End of Life (EOL) and End of Support (EOS) are as follows:

End of Life (EOL): The point in time in the life cycle of a product when the product is no longer being sold after it has reached the end of its useful life as defined by the manufacturer. For products that have reached EOL, a formal EOL process (notification to users, etc.) is carried out.

End of Support (EOS): The point in time in the life cycle of a product when the manufacturer discontinues all support activities during the product life cycle. Service support will not be provided after the end of support.

OEL has determined that the EOL of this product is 22<sup>nd</sup> October 2033.

\*10 years after the expiration date of the MDD EC Certificate has been set as the time when manufacturing and sales will end.

OEL has determined that the EOS of this product is 22<sup>nd</sup> October 2041.

\* EOL has been set in consideration of the product's shelf life of 8 years.

Yours Sincerely,

Toshiyuki Mori

Director, Quality Assurance Department

Otsuka Electronics Co.,Ltd.

Manufacturer : Otsuka Electronics Co., Ltd.

Address : 3-26-3, Shodai-Tajika, Hirakata-shi, Osaka, 573-1132, Japan